



Scan Tronic ApS
Sverigesvej 14
8700 Horsens, Danmark
CVR/VAT nr.: DK 25015886
Phone: +45 21 66 00 85
info@scan-tronic.dk
www.scan-tronic.dk

Calibration certificate

Oxygen Control **OC2016 nr: XXXX** with **Circuit board nr. XXX**, is calibrated with sensor **OS2014 nr: XXXX**. Calibrations points are 20,9% O₂ (atmospheric air) and calibration gas 1 % O₂ in Nitrogen 99% N₂.

Horsens, X.

CTO
B.Sc. EE Damir Josipovic